

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4503

Name: RITTER2  
Sterilizer ID: V2313228  
04/10/2025 02:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 101.1 F 0.0  
2:00 100.5 F 0.6  
4:00 109.7 F 2.6  
6:00 149.9 F 5.6  
8:00 172.7 F 9.2  
10:00 192.7 F 13.1  
12:00 217.7 F 5.1  
14:00 223.9 F 8.8  
16:00 235.0 F 12.9  
18:00 234.4 F 8.8  
20:00 244.1 F 12.4  
22:00 243.0 F 12.0  
24:00 254.1 F 17.8  
26:00 264.0 F 23.9  
28:00 270.8 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.6  
1:00 271.9 F 29.7  
2:00 271.8 F 29.8  
3:00 271.7 F 29.7  
4:00 271.6 F 29.6  
  
Min 271.6 F 29.5  
Max 273.2 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 232.8 F 0.0  
10:00 214.3 F 0.0  
15:00 219.5 F 0.0  
20:00 175.8 F 0.0  
25:00 169.5 F 0.0  
30:00 169.2 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 28:45  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:07:36  
CYCLE COMPLETE

04/10/2025 03:18 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_