

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3884

Name: RITTER 1  
Sterilizer ID: V2178198  
10/10/2023 03:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.2 F 0.0  
2:00 174.3 F 3.8  
4:00 196.8 F 8.2  
6:00 217.4 F 10.3  
8:00 231.1 F 8.1  
10:00 238.2 F 13.3  
12:00 238.7 F 10.7  
14:00 236.6 F 8.9  
16:00 250.8 F 15.7  
18:00 262.7 F 22.5  
20:00 271.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 271.3 F 28.5  
2:00 271.4 F 28.5  
3:00 271.8 F 28.5  
4:00 272.3 F 28.7  
  
Min 271.3 F 28.4  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 245.1 F 0.0  
10:00 198.6 F 0.0  
15:00 220.3 F 0.0  
20:00 192.1 F 0.0  
25:00 187.7 F 0.0  
30:00 186.3 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 20:03  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 00:59:16  
CYCLE COMPLETE

10/10/2023 04:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_