
Midmark M11 - v1.0.0
Total Cycles: 1549

Name: RITTER 1
Sterilizer ID: V2178198
01/20/2022 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.9 F 0.0
2:00 200.6 F 4.4
4:00 208.5 F 9.9
6:00 223.5 F 5.4
8:00 235.7 F 11.2
10:00 235.3 F 8.6
12:00 240.6 F 10.0
14:00 248.2 F 14.1
16:00 260.7 F 21.1
18:00 272.1 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.3 F 28.5
2:00 272.0 F 28.6
3:00 272.2 F 28.4
4:00 271.9 F 28.5

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 226.6 F 0.0
10:00 234.5 F 0.0
15:00 227.5 F 0.0
20:00 227.9 F 0.0
25:00 230.5 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 18:29
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

01/20/2022 04:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____