
Midmark M11 - v1.0.5
Total Cycles: 1389

Name: RITTER2
Sterilizer ID: V2313228
08/15/2022 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.5 F 0.0
2:00 127.8 F 0.0
4:00 155.7 F 0.0
6:00 190.5 F 0.4
8:00 196.9 F 4.0
10:00 218.1 F 8.4
12:00 229.5 F 13.0
14:00 231.0 F 8.7
16:00 241.0 F 12.3
18:00 241.1 F 11.5
20:00 238.6 F 10.0
22:00 251.9 F 16.5
24:00 263.2 F 23.1
26:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.3
2:00 271.3 F 28.6
3:00 272.5 F 29.4
4:00 271.6 F 28.6

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 216.8 F 0.0
10:00 187.2 F 0.0
15:00 184.1 F 0.0
20:00 182.1 F 0.0
25:00 182.4 F 0.0
30:00 181.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:15
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:04:16
CYCLE COMPLETE

08/15/2022 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____