
Midmark M11 - v1.0.0
Total Cycles: 2235

Name: RITTER 1
Sterilizer ID: V2178198
08/23/2022 11:06 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.7 F 0.0
2:00 144.0 F 1.4
4:00 158.8 F 4.8
6:00 180.4 F 9.2
8:00 202.1 F 5.7
10:00 222.8 F 7.5
12:00 236.3 F 12.8
14:00 236.7 F 9.6
16:00 236.9 F 8.5
18:00 248.3 F 14.2
20:00 260.6 F 21.1
22:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 271.4 F 28.0
2:00 271.6 F 28.0
3:00 271.9 F 28.1

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 229.8 F 0.0
Dry Cycle Aborted

FILLING: 1:07
HEATING: 22:27
STERILIZING: 3:00
VENTING: 2:51
DRYING: 0:19
TOTAL CYCLE: 00:29:44

08/23/2022 11:35 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____