
Midmark M11 - v1.0.5
Total Cycles: 2222

Name: RITTER2
Sterilizer ID: V2313228
04/10/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.8 F 0.0
2:00 200.3 F 4.1
4:00 220.8 F 9.3
6:00 226.1 F 7.2
8:00 234.1 F 10.5
10:00 232.4 F 7.9
12:00 240.5 F 10.9
14:00 247.0 F 14.0
16:00 259.7 F 20.9
18:00 271.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.1 F 29.2
2:00 271.4 F 29.1
3:00 272.0 F 29.2
4:00 271.2 F 28.9

Min 271.2 F 28.7
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 231.7 F 0.0
10:00 228.1 F 0.0
15:00 230.0 F 0.0
20:00 226.6 F 0.0
25:00 228.5 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

04/10/2023 12:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____