
Midmark M11 - v1.0.5
Total Cycles: 4168

Name: RITTER2
Sterilizer ID: V2313228
11/29/2024 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.4 F 0.0
2:00 164.5 F 0.2
4:00 169.9 F 2.3
6:00 197.8 F 6.9
8:00 224.7 F 11.6
10:00 226.1 F 7.8
12:00 235.9 F 13.4
14:00 240.2 F 11.7
16:00 239.8 F 10.9
18:00 253.7 F 17.8
20:00 265.8 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.6 F 29.8
2:00 271.4 F 29.0
3:00 272.2 F 29.9
4:00 272.2 F 29.4

Min 271.0 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 240.7 F 0.0
10:00 223.5 F 0.0
15:00 226.2 F 0.0
20:00 223.9 F 0.0
25:00 216.8 F 0.0
30:00 210.6 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:42
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

11/29/2024 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____