

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2446

Name: RITTER 1  
Sterilizer ID: V2178198  
10/18/2022 08:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.8 F 0.0  
2:00 138.5 F 1.5  
4:00 146.9 F 4.3  
6:00 173.0 F 8.2  
8:00 194.1 F 12.5  
10:00 220.8 F 5.7  
12:00 231.9 F 10.1  
14:00 233.4 F 8.4  
16:00 239.7 F 11.0  
18:00 234.5 F 8.3  
20:00 248.5 F 14.6  
22:00 259.7 F 20.8  
24:00 270.5 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.7 F 29.3  
2:00 271.8 F 29.3  
3:00 272.5 F 29.1  
4:00 271.8 F 29.3

Min 271.4 F 28.7  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
5:00 230.6 F 0.0  
10:00 203.7 F 0.0  
15:00 233.3 F 0.0  
20:00 240.0 F 0.0  
25:00 217.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:06  
HEATING: 25:32  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 27:54  
TOTAL CYCLE: 01:01:47

10/18/2022 09:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_