
Midmark M11 - v1.0.5
Total Cycles: 2811

Name: RITTER2
Sterilizer ID: V2313228
08/29/2023 12:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 166.8 F 0.1
4:00 194.8 F 3.8
6:00 212.7 F 8.6
8:00 227.8 F 10.9
10:00 233.2 F 10.0
12:00 234.7 F 8.9
14:00 242.1 F 11.9
16:00 246.0 F 13.6
18:00 258.0 F 20.2
20:00 269.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.6 F 29.7
2:00 271.7 F 28.9
3:00 272.1 F 29.7
4:00 272.0 F 29.2

Min 271.2 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 212.4 F 0.0
10:00 215.1 F 0.0
15:00 213.9 F 0.0
20:00 211.3 F 0.0
25:00 205.3 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:60
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:41
CYCLE COMPLETE

08/29/2023 01:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____