
Midmark M11 - v1.0.5
Total Cycles: 875

Name: RITTER2
Sterilizer ID: V2313228
03/14/2022 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.1 F 0.0
2:00 146.2 F 1.6
4:00 151.1 F 4.6
6:00 178.8 F 8.9
8:00 198.3 F 8.5
10:00 216.0 F 7.8
12:00 230.6 F 13.1
14:00 236.3 F 10.4
16:00 235.1 F 8.6
18:00 249.6 F 15.6
20:00 261.6 F 22.5
22:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.7
2:00 271.2 F 29.0
3:00 272.2 F 29.8
4:00 271.9 F 29.3

Min 271.2 F 29.0
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 231.2 F 0.0
10:00 223.1 F 0.0
15:00 225.5 F 0.0
20:00 223.4 F 0.0
25:00 224.3 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:31
CYCLE COMPLETE

03/14/2022 09:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____