

-----  
Midmark M11 - v1.0.5  
Total Cycles: 992

Name: RITTER2  
Sterilizer ID: V2313228  
04/14/2022 03:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.9 F 0.0  
2:00 193.8 F 1.7  
4:00 204.0 F 7.8  
6:00 221.4 F 11.7  
8:00 234.1 F 10.8  
10:00 235.2 F 9.4  
12:00 237.3 F 9.1  
14:00 251.4 F 16.6  
16:00 264.5 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.3 F 29.6  
2:00 271.3 F 29.2  
3:00 272.4 F 29.6  
4:00 271.4 F 28.9  
  
Min 271.2 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 217.6 F 0.0  
5:00 211.8 F 0.0  
10:00 217.6 F 0.0  
15:00 214.5 F 0.0  
20:00 214.5 F 0.0  
25:00 214.0 F 0.0  
30:00 215.4 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 17:47  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:56:36  
CYCLE COMPLETE

04/14/2022 04:04 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_