
Midmark M11 - v1.0.5
Total Cycles: 359

Name: RITTER2
Sterilizer ID: V2313228
09/14/2021 08:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.2 F 0.0
2:00 110.4 F 0.8
4:00 135.2 F 3.2
6:00 161.9 F 6.9
8:00 188.6 F 11.4
10:00 215.5 F 4.4
12:00 228.5 F 8.9
14:00 233.2 F 12.2
16:00 238.0 F 10.6
18:00 235.0 F 8.5
20:00 248.2 F 15.0
22:00 259.9 F 21.3
24:00 270.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 271.7 F 29.0
2:00 272.3 F 29.5
3:00 271.1 F 28.9
4:00 271.9 F 29.3

Min 271.0 F 28.7
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 233.8 F 0.0
10:00 198.0 F 0.0
15:00 226.9 F 0.0
20:00 217.5 F 0.0
25:00 216.4 F 0.0
30:00 164.6 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:49
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:02:49
CYCLE COMPLETE

09/14/2021 09:23 AM
ACCEPTED: X
REJECTED:
Operator: U01