
Midmark M11 - v1.0.0
Total Cycles: 4342

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2024 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.1 F 0.0
2:00 89.5 F 0.5
4:00 94.0 F 2.1
6:00 151.0 F 4.1
8:00 177.4 F 6.9
10:00 188.4 F 10.4
12:00 203.3 F 8.2
14:00 217.6 F 5.7
16:00 224.3 F 9.2
18:00 233.7 F 13.2
20:00 233.0 F 9.2
22:00 240.5 F 11.2
24:00 242.0 F 12.1
26:00 252.8 F 17.6
28:00 262.9 F 23.5
30:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.3 F 29.7
2:00 271.2 F 29.3
3:00 272.3 F 29.6
4:00 271.1 F 29.3

Min 271.1 F 29.1
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 222.6 F 0.0
10:00 240.1 F 0.0
15:00 228.9 F 0.0
20:00 218.8 F 0.0
25:00 225.1 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:24
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:09:07
CYCLE COMPLETE

02/28/2024 03:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____