
Midmark M11 - v1.0.5
Total Cycles: 1572

Name: RITTER2
Sterilizer ID: V2313228
10/06/2022 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.1 F 0.0
2:00 136.8 F 0.0
4:00 150.0 F 0.0
6:00 182.8 F 0.9
8:00 205.0 F 5.2
10:00 221.6 F 10.3
12:00 228.7 F 7.0
14:00 238.1 F 11.1
16:00 234.7 F 8.8
18:00 240.2 F 10.2
20:00 247.7 F 14.4
22:00 260.0 F 21.2
24:00 271.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.4 F 29.2
2:00 272.2 F 29.4
3:00 272.2 F 28.9
4:00 272.7 F 29.6

Min 271.7 F 28.7
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 206.2 F 0.0
10:00 205.4 F 0.0
15:00 216.0 F 0.0
20:00 214.5 F 0.0
25:00 243.5 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:02:53
CYCLE COMPLETE

10/06/2022 09:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____