
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006091
8:34 10/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:33
STERILIZING 3:37
136.4 C 222kPa 3:37
136.0 C 220kPa 4:07
136.4 C 224kPa 4:37
136.4 C 222kPa 5:07
136.4 C 224kPa 5:37
136.4 C 224kPa 6:07
136.3 C 222kPa 6:37
136.4 C 224kPa 7:07
136.4 C 224kPa 7:07
VENTING 7:08
AIR DRYING 7:59
CYCLE COMPLETE 14:10

Digital Signature #
D76512142840B784
