
Midmark M11 - v1.0.0
Total Cycles: 3037

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2023 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.5 F 0.0
2:00 196.9 F 4.4
4:00 203.3 F 9.4
6:00 228.9 F 7.3
8:00 233.3 F 10.4
10:00 232.9 F 7.3
12:00 244.4 F 12.4
14:00 245.0 F 12.6
16:00 257.9 F 19.5
18:00 269.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.5 F 28.9
2:00 271.6 F 28.2
3:00 271.6 F 28.7
4:00 272.4 F 28.7

Min 271.0 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 224.4 F 0.0
10:00 226.5 F 0.0
15:00 235.8 F 0.0
20:00 225.5 F 0.0
25:00 227.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

03/27/2023 12:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____