

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5386

Name: RITTER 1  
Sterilizer ID: V2178198  
03/05/2025 09:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.6 F 0.0  
2:00 77.8 F 0.3  
4:00 88.6 F 1.5  
6:00 128.1 F 4.2  
8:00 152.2 F 7.2  
10:00 171.8 F 10.9  
12:00 190.2 F 3.1  
14:00 206.1 F 4.9  
16:00 217.3 F 8.2  
18:00 228.2 F 11.9  
20:00 227.7 F 6.8  
22:00 240.0 F 11.8  
24:00 236.6 F 9.2  
26:00 249.7 F 15.4  
28:00 260.9 F 21.7  
30:00 271.1 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.8 F 29.2  
2:00 271.8 F 28.7  
3:00 272.8 F 29.2  
4:00 271.8 F 28.6

Min 271.6 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 191.3 F 0.0  
10:00 199.2 F 0.0  
15:00 197.4 F 0.0  
20:00 200.9 F 0.0  
25:00 200.3 F 0.0  
30:00 199.4 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 30:43  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:08:56  
CYCLE COMPLETE

03/05/2025 10:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_