
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000684
8:23 15/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:14
STERILIZING 3:16
136.0 C 227kPa 3:16
135.9 C 228kPa 3:46
136.3 C 228kPa 4:16
136.4 C 233kPa 4:46
136.2 C 232kPa 5:16
136.4 C 233kPa 5:46
136.3 C 232kPa 6:16
136.4 C 231kPa 6:46
136.4 C 231kPa 6:46
VENTING 6:46
AIR DRYING 7:40
CYCLE COMPLETE 7:41

Digital Signature #
D21F10B305EE460F
