

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002209  
12:25 04/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:02  
STERILIZING 4:14  
136.4 C 224kPa 4:14  
136.3 C 226kPa 4:44  
136.4 C 224kPa 5:14  
136.1 C 223kPa 5:44  
136.2 C 224kPa 6:14  
136.4 C 225kPa 6:44  
136.3 C 225kPa 7:14  
136.2 C 225kPa 7:44  
136.2 C 225kPa 7:44  
VENTING 7:44  
AIR DRYING 8:39  
CYCLE COMPLETE 8:47  
-----

Digital Signature #  
DB77BC6258848C2F  
-----