
Midmark M11 - v1.0.0
Total Cycles: 3820

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2023 12:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.0 F 0.0
2:00 178.9 F 2.7
4:00 199.3 F 7.3
6:00 215.3 F 12.3
8:00 225.5 F 7.5
10:00 238.5 F 12.5
12:00 237.9 F 10.0
14:00 239.0 F 8.8
16:00 249.7 F 14.9
18:00 261.1 F 21.3
20:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.6 F 28.8
2:00 271.3 F 28.1
3:00 272.2 F 29.0
4:00 272.1 F 28.4

Min 271.1 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 235.2 F 0.0
10:00 222.0 F 0.0
15:00 216.4 F 0.0
20:00 221.6 F 0.0
25:00 225.5 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

09/26/2023 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____