
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003312
12:21 28/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:50
STERILIZING 3:59
136.5 C 222kPa 3:59
136.2 C 223kPa 4:29
136.4 C 224kPa 4:59
136.2 C 219kPa 5:29
136.4 C 224kPa 5:59
136.1 C 221kPa 6:29
136.4 C 224kPa 6:59
136.4 C 225kPa 7:29
136.4 C 225kPa 7:29
VENTING 7:29
AIR DRYING 8:25
CYCLE COMPLETE 8:46

Digital Signature #
F3B77102A42601D2
