
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000745
12:16 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:14
STERILIZING 3:19
136.4 C 232kPa 3:19
136.4 C 233kPa 3:49
136.4 C 231kPa 4:19
136.4 C 231kPa 4:49
136.2 C 230kPa 5:19
136.4 C 232kPa 5:49
136.4 C 233kPa 6:19
136.4 C 233kPa 6:49
136.4 C 233kPa 6:49
VENTING 6:49
AIR DRYING 7:41
CYCLE COMPLETE 8:07

Digital Signature #
3DE732F12FA0CCE9
