
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002647
8:02 03/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:13
STERILIZING 4:21
136.0 C 220kPa 4:21
136.1 C 222kPa 4:51
136.0 C 221kPa 5:21
136.0 C 221kPa 5:51
136.2 C 222kPa 6:21
136.4 C 224kPa 6:51
136.3 C 222kPa 7:21
136.4 C 223kPa 7:51
136.4 C 223kPa 7:51
VENTING 7:51
AIR DRYING 8:43
CYCLE COMPLETE 8:45

Digital Signature #
10453230C4733DD5
