
Midmark M11 - v1.0.5
Total Cycles: 743

Name: RITTER2
Sterilizer ID: V2313228
02/01/2022 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 172.0 F 2.8
4:00 194.6 F 7.9
6:00 215.9 F 13.4
8:00 229.1 F 8.4
10:00 238.9 F 11.0
12:00 243.9 F 13.1
14:00 244.6 F 12.7
16:00 258.0 F 19.9
18:00 269.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.8 F 29.5
2:00 271.8 F 28.7
3:00 272.0 F 29.4
4:00 272.6 F 29.4

Min 271.3 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 222.3 F 0.0
10:00 228.6 F 0.0
15:00 222.7 F 0.0
20:00 220.3 F 0.0
25:00 218.0 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

02/01/2022 11:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____