
Midmark M11 - v1.0.0
Total Cycles: 2980

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2023 12:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.8 F 0.0
2:00 187.3 F 4.0
4:00 205.1 F 9.0
6:00 222.1 F 7.5
8:00 232.1 F 10.1
10:00 234.8 F 7.8
12:00 243.1 F 11.7
14:00 246.1 F 13.1
16:00 259.2 F 20.3
18:00 270.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.4 F 28.7
2:00 271.1 F 28.2
3:00 272.4 F 28.9
4:00 271.6 F 28.2

Min 271.1 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 238.3 F 0.0
10:00 235.1 F 0.0
15:00 229.2 F 0.0
20:00 229.5 F 0.0
25:00 220.1 F 0.0
30:00 215.2 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

03/08/2023 01:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____