

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000087  
10:48 13/02/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:32  
PRESSURIZING 3:33  
STERILIZING 4:51  
136.5 C 223kPa 4:51  
136.2 C 222kPa 5:21  
136.4 C 224kPa 5:51  
136.3 C 222kPa 6:21  
136.4 C 224kPa 6:51  
CYCLE ABORTED 7:01  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 7:01  
-----

Digital Signature #  
5A99D014CBA166F8  
-----