

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2269

Name: RITTER 1  
Sterilizer ID: V2178198  
08/30/2022 04:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.8 F 0.0  
2:00 187.3 F 0.8  
4:00 208.7 F 3.5  
6:00 212.1 F 8.5  
8:00 226.3 F 13.1  
10:00 232.1 F 8.2  
12:00 242.7 F 13.2  
14:00 240.7 F 11.0  
16:00 238.0 F 9.5  
18:00 252.2 F 16.3  
20:00 263.7 F 23.1  
22:00 271.3 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.5 F 29.1  
2:00 271.8 F 28.9  
3:00 272.4 F 29.0  
4:00 271.5 F 28.8

Min 271.5 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.6 F 0.0  
5:00 237.9 F 0.0  
10:00 228.5 F 0.0  
15:00 229.9 F 0.0  
20:00 224.8 F 0.0  
25:00 232.2 F 0.0  
30:00 217.5 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 22:48  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:01:38  
CYCLE COMPLETE

08/30/2022 05:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_