

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 008037  
13:39 07/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:40  
STERILIZING 3:39  
136.0 C 220kPa 3:39  
136.1 C 220kPa 4:09  
136.1 C 222kPa 4:39  
136.4 C 222kPa 5:09  
136.4 C 222kPa 5:39  
136.2 C 222kPa 6:09  
136.4 C 224kPa 6:39  
136.4 C 221kPa 7:09  
136.4 C 221kPa 7:09  
VENTING 7:09  
AIR DRYING 8:04  
CYCLE COMPLETE 38:05

-----  
Digital Signature #  
9DA1DA5798B54212  
-----