

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001672  
16:47 03/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:11  
PRESSURIZING 2:15  
STERILIZING 3:13  
136.5 C 225kPa 3:13  
136.2 C 224kPa 3:43  
136.2 C 223kPa 4:13  
136.4 C 226kPa 4:43  
136.4 C 224kPa 5:13  
136.4 C 226kPa 5:43  
136.4 C 226kPa 6:13  
136.4 C 226kPa 6:43  
136.4 C 226kPa 6:43  
VENTING 6:44  
AIR DRYING 7:34  
CYCLE COMPLETE 8:00  
-----

Digital Signature #  
B8807342AF05EADF  
-----