
Midmark M11 - v1.0.0
Total Cycles: 1679

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2022 01:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 195.4 F 3.4
4:00 201.1 F 8.3
6:00 220.3 F 8.8
8:00 229.9 F 9.1
10:00 237.4 F 9.7
12:00 243.1 F 12.2
14:00 241.3 F 10.8
16:00 254.6 F 17.5
18:00 266.0 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.6 F 28.1
3:00 272.0 F 28.7
4:00 272.3 F 28.7

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 222.6 F 0.0
10:00 226.1 F 0.0
15:00 230.1 F 0.0
20:00 228.4 F 0.0
25:00 228.3 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 19:41
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

02/28/2022 02:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____