
Midmark M11 - v1.0.5
Total Cycles: 1163

Name: RITTER2
Sterilizer ID: V2313228
06/07/2022 06:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.6 F 0.0
2:00 76.0 F 0.0
4:00 79.2 F 0.0
6:00 103.5 F 0.1
8:00 139.6 F 2.8
10:00 166.7 F 6.0
12:00 186.6 F 9.5
14:00 210.3 F 13.4
16:00 230.4 F 7.2
18:00 238.2 F 11.4
20:00 232.9 F 7.7
22:00 245.3 F 13.2
24:00 242.4 F 11.6
26:00 254.2 F 17.7
28:00 264.5 F 23.9
30:00 270.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.9 F 29.5
2:00 272.0 F 29.8
3:00 271.9 F 29.2
4:00 272.6 F 29.9

Min 271.3 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 245.1 F 0.0
10:00 225.7 F 0.0
15:00 229.5 F 0.0
20:00 224.8 F 0.0
25:00 233.1 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:54
CYCLE COMPLETE

06/07/2022 07:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____