
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007116
13:17 26/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:01
STERILIZING 4:00
136.4 C 223kPa 4:00
136.1 C 223kPa 4:30
136.4 C 225kPa 5:00
136.4 C 222kPa 5:30
136.2 C 222kPa 6:00
136.3 C 223kPa 6:30
136.4 C 223kPa 7:00
136.3 C 221kPa 7:30
136.3 C 221kPa 7:30
VENTING 7:30
AIR DRYING 8:25
CYCLE COMPLETE 8:30

Digital Signature #
32749C865E2742D4
