
Midmark M11 - v1.0.0
Total Cycles: 3830

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2023 02:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.1 F 0.0
2:00 168.3 F 0.1
4:00 188.1 F 3.5
6:00 209.9 F 8.2
8:00 223.3 F 13.2
10:00 228.7 F 8.5
12:00 240.1 F 12.7
14:00 240.5 F 11.2
16:00 238.3 F 9.5
18:00 252.4 F 16.3
20:00 263.9 F 23.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.0 F 28.1
2:00 272.5 F 28.6
3:00 272.2 F 28.7
4:00 271.7 F 28.0

Min 271.5 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 232.5 F 0.0
10:00 227.6 F 0.0
15:00 228.7 F 0.0
20:00 231.3 F 0.0
25:00 228.2 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

09/27/2023 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____