

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1335

Name: RITTER2  
Sterilizer ID: V2313228  
07/25/2022 08:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.9 F 0.0  
2:00 171.2 F 0.0  
4:00 173.6 F 0.7  
6:00 195.2 F 5.2  
8:00 220.8 F 10.2  
10:00 225.8 F 7.6  
12:00 236.9 F 11.9  
14:00 235.9 F 9.7  
16:00 238.9 F 9.2  
18:00 247.9 F 15.0  
20:00 259.8 F 21.7  
22:00 270.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.4 F 29.9  
2:00 271.3 F 29.3  
3:00 272.5 F 30.1  
4:00 271.9 F 29.5

Min 271.3 F 29.2  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 234.5 F 0.0  
10:00 233.4 F 0.0  
15:00 229.7 F 0.0  
20:00 220.0 F 0.0  
25:00 227.6 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 22:47  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:49  
CYCLE COMPLETE

07/25/2022 09:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_