
Midmark M11 - v1.0.5
Total Cycles: 1997

Name: RITTER2
Sterilizer ID: V2313228
02/02/2023 08:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.2 F 0.0
2:00 78.7 F 0.0
4:00 87.4 F 0.0
6:00 122.3 F 0.7
8:00 158.8 F 3.7
10:00 179.5 F 7.0
12:00 200.7 F 10.6
14:00 213.5 F 5.2
16:00 224.3 F 8.7
18:00 235.1 F 13.1
20:00 236.1 F 9.9
22:00 237.1 F 9.1
24:00 246.9 F 14.1
26:00 258.0 F 20.3
28:00 268.5 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.2 F 29.6
2:00 271.3 F 29.2
3:00 272.1 F 29.5
4:00 271.1 F 29.0

Min 271.1 F 28.9
Max 272.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 201.0 F 0.0
10:00 201.6 F 0.0
15:00 207.3 F 0.0
20:00 207.6 F 0.0
25:00 209.4 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:12
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:07:08
CYCLE COMPLETE

02/02/2023 10:00 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____