
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001490
9:22 10/12/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:52
PRESSURIZING 6:54
STERILIZING 8:15
136.0 C 220kPa 8:15
136.5 C 219kPa 8:45
135.8 C 219kPa 9:15
136.2 C 220kPa 9:45
136.4 C 221kPa 10:15
136.3 C 221kPa 10:45
136.4 C 218kPa 11:15
136.4 C 219kPa 11:45
VENTING 11:45
AIR DRYING 12:47
CYCLE COMPLETE 42:47

Digital Signature #
875F03D842D1FB93
