

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4198

Name: RITTER2  
Sterilizer ID: V2313228  
12/10/2024 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.9 F 0.0  
2:00 69.1 F 0.0  
4:00 74.3 F 0.0  
6:00 85.5 F 0.0  
8:00 139.5 F 0.1  
10:00 160.1 F 2.5  
12:00 178.1 F 5.3  
14:00 189.0 F 8.3  
16:00 202.0 F 11.3  
18:00 219.7 F 5.7  
20:00 227.2 F 8.5  
22:00 235.8 F 12.6  
24:00 234.8 F 9.4  
26:00 240.1 F 10.7  
28:00 244.8 F 13.2  
30:00 255.6 F 18.9  
32:00 265.4 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.3 F 29.6  
2:00 271.6 F 29.4  
3:00 272.1 F 29.3  
4:00 271.8 F 29.7  
  
Min 271.4 F 28.9  
Max 272.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.9 F 0.0  
5:00 209.4 F 0.0  
10:00 208.2 F 0.0  
15:00 219.5 F 0.0  
20:00 204.6 F 0.0  
25:00 212.4 F 0.0  
30:00 210.5 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 33:55  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:12:31  
CYCLE COMPLETE

12/10/2024 11:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_