

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000858  
13:04 12/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:03  
PRESSURIZING 3:07  
STERILIZING 4:45  
136.4 C 230kPa 4:45  
136.4 C 232kPa 5:15  
136.4 C 232kPa 5:45  
136.1 C 230kPa 6:15  
136.1 C 229kPa 6:45  
136.3 C 231kPa 7:15  
136.4 C 232kPa 7:45  
136.2 C 230kPa 8:15  
136.2 C 230kPa 8:15  
VENTING 8:15  
AIR DRYING 9:08  
CYCLE COMPLETE 24:08  
-----

Digital Signature #  
EA8E0D2D82F8A629  
-----