

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4401

Name: RITTER2  
Sterilizer ID: V2313228  
03/05/2025 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.1 F 0.0  
2:00 119.6 F 1.2  
4:00 130.4 F 3.5  
6:00 164.6 F 7.1  
8:00 193.9 F 11.0  
10:00 213.2 F 5.4  
12:00 224.9 F 9.3  
14:00 231.4 F 10.4  
16:00 236.9 F 11.2  
18:00 235.7 F 9.5  
20:00 249.3 F 15.9  
22:00 260.8 F 22.4  
24:00 270.9 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.9  
1:00 272.7 F 30.7  
2:00 271.6 F 30.1  
3:00 272.3 F 30.4  
4:00 271.7 F 30.3

Min 271.0 F 29.7  
Max 273.5 F 31.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 227.7 F 0.0  
10:00 221.1 F 0.0  
15:00 228.3 F 0.0  
20:00 219.6 F 0.0  
25:00 201.3 F 0.0  
30:00 177.8 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:48  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:03:27  
CYCLE COMPLETE

03/05/2025 01:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_