
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004698
10:45 14/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:50
STERILIZING 3:55
136.4 C 224kPa 3:55
136.4 C 223kPa 4:25
136.4 C 223kPa 4:55
136.1 C 220kPa 5:25
136.2 C 221kPa 5:55
136.3 C 222kPa 6:25
136.4 C 222kPa 6:55
136.4 C 224kPa 7:25
136.4 C 224kPa 7:25
VENTING 7:25
AIR DRYING 8:16
CYCLE COMPLETE 8:27

Digital Signature #
4FC32CFAC9227F39
