
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000192
16:18 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:50
PRESSURIZING 0:55
STERILIZING 2:26
136.4 C 225kPa 2:26
136.3 C 224kPa 2:56
136.3 C 223kPa 3:26
136.4 C 224kPa 3:56
136.4 C 225kPa 4:26
136.4 C 225kPa 4:56
136.3 C 221kPa 5:26
136.3 C 222kPa 5:56
136.3 C 222kPa 5:56
VENTING 5:56
AIR DRYING 7:00
CYCLE COMPLETE 7:01

Digital Signature #
1CAF5BF19A76592D
