
Midmark M11 - v1.0.5
Total Cycles: 1408

Name: RITTER2
Sterilizer ID: V2313228
08/24/2022 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.5 F 0.0
2:00 146.2 F 1.2
4:00 161.0 F 3.5
6:00 197.6 F 7.6
8:00 214.0 F 12.6
10:00 225.8 F 7.5
12:00 238.0 F 12.6
14:00 236.2 F 10.1
16:00 238.1 F 8.8
18:00 248.7 F 15.1
20:00 260.6 F 21.8
22:00 271.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.4 F 29.6
2:00 271.7 F 29.5
3:00 272.4 F 29.5
4:00 271.6 F 29.4

Min 271.4 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 215.1 F 0.0
10:00 222.3 F 0.0
15:00 224.9 F 0.0
20:00 223.8 F 0.0
25:00 226.2 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

08/24/2022 02:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____