
Midmark M11 - v1.0.5
Total Cycles: 1258

Name: RITTER2
Sterilizer ID: V2313228
07/05/2022 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.3 F 0.0
2:00 144.5 F 1.8
4:00 167.6 F 5.3
6:00 188.3 F 9.5
8:00 209.0 F 8.2
10:00 222.5 F 8.3
12:00 233.0 F 13.3
14:00 236.0 F 10.1
16:00 235.6 F 8.7
18:00 246.9 F 14.4
20:00 258.4 F 20.7
22:00 268.7 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 272.8 F 30.0
1:00 272.6 F 30.3
2:00 272.1 F 30.6
3:00 271.7 F 30.0
4:00 272.5 F 30.3

Min 271.7 F 29.9
Max 272.8 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 197.0 F 0.0
10:00 212.8 F 0.0
15:00 220.9 F 0.0
20:00 224.4 F 0.0
25:00 223.9 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 23:56
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

07/05/2022 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____