
Midmark M11 - v1.0.0
Total Cycles: 5164

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2024 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.9 F 0.0
2:00 193.7 F 0.6
4:00 201.9 F 2.1
6:00 216.8 F 6.5
8:00 232.3 F 11.0
10:00 227.3 F 6.7
12:00 240.9 F 12.7
14:00 239.9 F 10.9
16:00 238.1 F 9.8
18:00 252.0 F 16.7
20:00 264.0 F 23.6
22:00 271.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.8 F 29.4
2:00 272.1 F 28.9
3:00 271.2 F 28.5
4:00 272.5 F 29.4

Min 271.2 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 249.4 F 0.0
10:00 226.9 F 0.0
15:00 232.8 F 0.0
20:00 210.9 F 0.0
25:00 210.3 F 0.0
30:00 199.5 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:01:03
CYCLE COMPLETE

12/05/2024 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____