
Midmark M11 - v1.0.5
Total Cycles: 3854

Name: RITTER2
Sterilizer ID: V2313228
08/13/2024 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.4 F 0.0
2:00 115.6 F 0.6
4:00 139.4 F 3.1
6:00 170.7 F 6.7
8:00 193.1 F 10.6
10:00 209.8 F 5.2
12:00 221.6 F 7.7
14:00 233.6 F 12.2
16:00 232.9 F 8.8
18:00 239.4 F 10.9
20:00 244.6 F 13.1
22:00 256.3 F 19.3
24:00 266.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 272.4 F 29.6
2:00 271.2 F 29.3
3:00 272.4 F 29.5
4:00 271.3 F 29.3

Min 271.0 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 222.3 F 0.0
10:00 221.2 F 0.0
15:00 226.1 F 0.0
20:00 232.6 F 0.0
25:00 230.0 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:34
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:04:04
CYCLE COMPLETE

08/13/2024 04:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____