
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000408
11:45 26/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:10
STERILIZING 4:14
136.0 C 224kPa 4:14
136.2 C 232kPa 4:44
136.4 C 233kPa 5:14
136.3 C 233kPa 5:44
136.4 C 232kPa 6:14
136.4 C 233kPa 6:44
136.2 C 231kPa 7:14
136.4 C 233kPa 7:44
136.4 C 233kPa 7:44
VENTING 7:44
AIR DRYING 8:40
CYCLE COMPLETE 8:41

Digital Signature #
1011A6CF1721BF5C
