
Midmark M11 - v1.0.5
Total Cycles: 51

Name: RITTER2
Sterilizer ID: V2313228
05/27/2021 01:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 185.9 F 3.8
4:00 197.1 F 9.1
6:00 213.8 F 7.5
8:00 229.3 F 9.7
10:00 234.9 F 9.2
12:00 241.8 F 11.8
14:00 246.7 F 13.9
16:00 258.9 F 20.7
18:00 270.0 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 271.4 F 29.0
2:00 272.2 F 29.7
3:00 272.6 F 29.9
4:00 272.6 F 29.6

Min 271.3 F 28.8
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 221.3 F 0.0
10:00 220.8 F 0.0
15:00 213.7 F 0.0
20:00 208.0 F 0.0
25:00 205.2 F 0.0
30:00 201.9 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

05/27/2021 02:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____