
Midmark M11 - v1.0.5
Total Cycles: 2853

Name: RITTER2
Sterilizer ID: V2313228
09/08/2023 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 170.5 F 2.9
4:00 188.6 F 7.2
6:00 207.2 F 12.1
8:00 220.2 F 7.0
10:00 231.6 F 12.0
12:00 233.9 F 9.4
14:00 238.1 F 9.8
16:00 246.7 F 14.2
18:00 258.1 F 20.5
20:00 268.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 271.6 F 29.2
2:00 273.3 F 30.3
3:00 272.4 F 30.5
4:00 271.7 F 29.8

Min 270.9 F 29.1
Max 273.4 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.9 F 0.0
5:00 197.7 F 0.0
10:00 194.6 F 0.0
15:00 193.9 F 0.0
20:00 190.9 F 0.0
25:00 188.1 F 0.0
30:00 185.4 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

09/08/2023 01:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____