
Midmark M11 - v1.0.5
Total Cycles: 2786

Name: RITTER2
Sterilizer ID: V2313228
08/17/2023 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 193.1 F 0.0
4:00 213.5 F 4.3
6:00 228.6 F 9.5
8:00 233.9 F 9.3
10:00 242.4 F 13.0
12:00 242.6 F 12.2
14:00 242.7 F 11.9
16:00 256.2 F 19.0
18:00 267.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.8 F 29.6
2:00 271.6 F 28.7
3:00 272.8 F 29.6
4:00 271.9 F 28.9

Min 271.5 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 211.9 F 0.0
10:00 215.4 F 0.0
15:00 214.0 F 0.0
20:00 211.9 F 0.0
25:00 210.3 F 0.0
30:00 208.4 F 0.0

Drying Complete

FILLING: 1:56
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

08/17/2023 03:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____