
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.9uS / 1.8ppm
CYCLE NUMBER 000042
6:53 24/08/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 4:58
STERILIZING 6:18
136.0 C 221kPa 6:18
136.4 C 226kPa 6:48
136.3 C 226kPa 7:18
135.9 C 223kPa 7:48
136.1 C 225kPa 8:18
136.4 C 227kPa 8:48
136.4 C 227kPa 9:18
136.2 C 226kPa 9:48
VENTING 9:48
AIR DRYING 10:42
CYCLE COMPLETE 40:42

Digital Signature #
D782737D6D25C3FA
